

FDA-RFQ-1181318 Amendment 0001  
Various Lab Equipment

<b>Document Type:</b>	Amendment 0001 Questions and Answers
<b>Solicitation Number:</b>	FDA-RFQ-1179320 Amend 0001
<b>Posted Date:</b>	April 2017
<b>Response Date:</b>	21 April 2017
<b>NAICS Code:</b>	334516, Analytical Laboratory Instrument Manufacturing, Small Size Standard 1,000 Employees.

**The purpose of this amendment is to: provide questions and answers for the RFQ.**

**Part 1 – QUESTIONS AND ANSWERS**

1. For all of the centrifuges(10008 , 10009, 10010), there's no mention of any rotors. Are those needed and if so, what rotors?

10008- We would like the 4x750ml and 14x50ml swinging rotor

10009 and 10010-It should be a 24x1.5ml rotor

2. For items 10001 & 10002 freezer do you need manual or auto defrost?

We would prefer auto-defrost but it is not necessary

3. For 10004, is the pH meter handheld (portable) or bench top?

This should be a benchtop

**Part 2 - RESPONSE DATES**

Response Date remains the same.

RFQ Due Date: All quotes and revisions shall be submitted via email to Christine Russman at Christine.Russman@fda.hhs.gov before 2:00 PM Eastern Time on 21 April 2017.

All other terms and conditions remain unchanged.